

# Bold

Aurelia® Bold® Black Nitrile  
Powder Free Examination Gloves

Free Sample  
Size Medium



**Durable Black Nitrile  
Maximum Strength**

- Robust average 6.0mil finger-5.0mil thickness palm for enhanced puncture & abrasion resistance
- Fully-textured finish for optimal gripping action
- 100% latex free, High contrast black color
- ASTM D6978-05 tested for Chemotherapy Drugs Permeation Resistance



## AURELIA® GLOVES

[www.aureliagloves.com](http://www.aureliagloves.com) | 1-877-AURELIA (1-877-287-3542)

# Bold

## Product description & specification

|                            |                                                                                                                                                                                                                                                                   |                                                                                                           |                                                                                              |
|----------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------|
| <b>Label</b>               | Aurelia Bold                                                                                                                                                                                                                                                      |                                                                                                           |                                                                                              |
| <b>Product No.</b>         | 73995 (XS) 73996 (S) 73997 (M) 73998 (L) 73998 (L) 73999 (XL) 73999 (XXL) 73990                                                                                                                                                                                   |                                                                                                           |                                                                                              |
| <b>FDA Code</b>            | LZA                                                                                                                                                                                                                                                               |                                                                                                           |                                                                                              |
| <b>Material</b>            | Proprietary blend Nitrile Butadiene Rubber (NBR)                                                                                                                                                                                                                  |                                                                                                           |                                                                                              |
| <b>Design</b>              | Beaded cuff, Powder-free & Ambidextrous                                                                                                                                                                                                                           |                                                                                                           |                                                                                              |
| <b>Color</b>               | Satin Black                                                                                                                                                                                                                                                       |                                                                                                           |                                                                                              |
| <b>Physical Dimension</b>  | Dimension – Palm Width (in)                                                                                                                                                                                                                                       | <b>(ASTM 6319)</b><br>Size S : 3 ± 0.4<br>Size M : 3.7 ± 0.4<br>Size L : 4.3 ± 0.4<br>Size XL : 4.5 ± 0.4 | <b>Bold</b><br>Size S : 3.15 avg<br>Size M : 3.75 avg<br>Size L : 4.3 avg<br>Size XL : ≥ 4.7 |
|                            | Dimension – Gloves Length (in)                                                                                                                                                                                                                                    | Size S : 8.7 min<br>Size M : 8.7 min<br>Size L : 8.7 min<br>Size XL : 8.7 min                             | Size S : ≥9.4<br>Size M : ≥9.4<br>Size L : ≥9.4<br>Size XL : ≥9.4                            |
| <b>Physical Properties</b> | Thickness Single Wall (mm)                                                                                                                                                                                                                                        | Finger : 6mil(avg)<br>Palm : 5mil(avg)                                                                    |                                                                                              |
|                            | ASTM D6319<br>Tensile Strength (MPa)<br>Ultimate Elongation (%)                                                                                                                                                                                                   | Unaged : 14 Min<br>Unaged : 500 Min                                                                       | Aged : 14 Min<br>Aged : 400 Min                                                              |
|                            | EN 455<br>Force at Break (N)                                                                                                                                                                                                                                      | Unaged:<br>Median ≥6N                                                                                     | Aged:<br>Median ≥6N                                                                          |
| <b>Certification</b>       | FDA, CE,                                                                                                                                                                                                                                                          |                                                                                                           |                                                                                              |
|                            | ISO 13485:2016                                                                                                                                                                                                                                                    |                                                                                                           |                                                                                              |
| <b>Product Test</b>        | EN455, ASTM including ASTM D6978-05, NZ/AU4011.1:2014, ISO 11193-1:2008                                                                                                                                                                                           |                                                                                                           |                                                                                              |
| <b>Country of Origin</b>   | <b><u>Made in Malaysia</u></b>                                                                                                                                                                                                                                    |                                                                                                           |                                                                                              |
| <b>Intended Usage</b>      | Nitrile is waterproof, greaseproof, oil proof and most importantly, chemical resistant to a range of common chemicals and substances. Nitrile gloves are also commonly used in the medical industry due to their high puncture resistance, and dexterous strength |                                                                                                           |                                                                                              |

## CONFORMITY

- The general safety and performance requirements of FDA Medical Device Regulation for Class 1 medical devices (21 CFR 880.6250) and complies with all general controls (section 513(a)(1)(A) of the Federal Food, Drug, and Cosmetic Act (Act); 21 U.S.C. 360c(a)(1)(A)), which include: Establishment Registration Device Listing Premarket Notification 510(k) Labeling Quality System and Medical Device Reporting (MDR)
- The general safety and performance requirements of Annex I, Medical Device Regulation (EU) 2017/745 for Class I medical devices. Classification: Class I based on Rule 5 transient use, Annex VIII of the Medical Device Regulation (EU) 2017/745 - With the national standard transposing harmonized standard EN455-1, EN455-2, EN455-3 and EN455-4 and is self-certified as a Class I non-sterile medical device
- The standard specification requirements for ASTM D6319- Freedom From Holes, Physical Dimensions Test, Physical Requirement Test and Packaging and ASTM F1671-07 – Phi-x174 Bacteriophage Penetration Test
- EEC regulations concerning the conformity of materials and products that are allowed to come into contact with food. In accordance with Regulation EEC 1935/2004, Regulation EC 10/2011 & Regulation (EC) No 2023/2006. The gloves are suitable for contact with dry, fatty, alcoholic and aqueous food for short term contact.
- The gloves are certified manufactured according to ISO 9001:2015 and ISO 13485:2016 Quality Management Systems

Manufactured by,  
Supermax Healthcare Inc  
1899 Sequoia Dr  
Aurora, IL 60506

Factory Location:  
Supermax Corporation  
Malaysia

1 877-287-3542  
[info@aureliagloves.com](mailto:info@aureliagloves.com)

