



Operation Manual

Heating & Cooling block

Model : CCB-350

Manual No. : J5011L002 Version : 2.1



WARNING

Please read this user manual carefully before using the product. Users should operate the product according to the operating instructions contained in this user manual. Users should be aware of cautions and warnings for your safety. Failure to follow these guidelines may result in serious injury or damage to your property.

Thank you for purchasing JEIOTECH's products.

JEIOTECH Co., Ltd. is committed to offer the customer service including product sales, Technical service, and follow-up management. If you have any question about this user manual, please contact our sales department.

Quality Management System ISO 9001



ISO 9001 is a standard that defines the requirements for a Quality Management System (QMS). It ensures the suitability of implementation in order for suppliers to continuously provide products and services that meet the buyer's requirements. JEIOTECH Co., Ltd. obtained ISO 9001 certification from *Korean Foundation for Quality (KFQ)* since 1998 to provide better service. Under this certification system, we design, develop, produce and sell a wide range of products related to chemical engineering, environment and life. We also provide reliability testing equipment for electrical and electronic engineering. (Dryer, incubator, constant temperature and humidity chamber, constant temperature water tank, low temperature circulation tank, heat exchanger, stirrer, etc.)

You can download the above image by visiting the following JEIOTECH site.

<http://www.jeitech.com>

Disclaimer

The material in this manual is for the users' information only. The products it describes are subject to change without prior notice, due to the manufacturer's continuous development program.

Copyright

No part of this manual may be reproduced or transmitted in any form or by any means, including photocopying, recording, or using information storage and retrieval systems, for any purpose other than the purchaser's own use, without the express written permission of JEIOTECH Co., Ltd.

©2026. All Rights Reserved. JEIOTECH Co., Ltd.

Third-party product names or brand names mentioned in this user manual may be the trademarks of their respective owners. JEIOTECH has no connection with those companies and we do not endorse or recommend them directly nor indirectly.

Table of contents

1.0	Safety	1
1.1	Operation Manual.....	1
1.2	WARNING/CAUTION/NOTICE Alerts	1
1.3	Exemption for Responsibility	2
1.4	Warning Statement.....	2
1.5	Caution Statement	3
2.0	Functional Description	4
2.1	Introduction	4
2.2	Feature	4
2.2.1	Excellent Performance	4
2.2.2	Safety	4
2.2.3	Ease of Use	5
2.3	Construction	6
3.0	Installation	7
3.1	Unpacking Package	7
3.2	Checking Instrument Components	7
3.3	Installing Environment.....	8
3.4	Location Conditions.....	8
3.5	Checking Points	8
3.6	Connecting to Main Power Supply.....	9
3.7	Instrument Start and Stop.....	9
4.0	Operation	11
4.1	Control Panel and Function	11
4.2	General Modes.....	12
4.2.1	Setting Temperature	12
4.3	Timer Modes.....	15
4.3.1	Setting T1 Timer.....	15
4.3.2	Setting T2 Timer.....	17
4.3.3	Checking Remaining Time during Timer Operation	19
4.3.4	Stopping Timer during Operation.....	19
4.4	Program Mode	20
4.4.1	Starting User Program	21
4.4.2	Display during Program Operation	22
4.4.3	Editing User Program	24
4.4.4	Viewing Program Setting.....	27
4.4.5	Stopping Program	29
4.4.6	Confirming Program End.....	29
4.4.7	Resetting Program	30
4.5	How to Replace a Block	32

4.6	Offset.....	3 3
5.0	Safety Device.....	3 5
6.0	Maintenance	3 6
6.1	Periodic Maintenance	3 6
6.2	Cleaning Product.....	3 6
6.2.1	Main Body.....	3 6
6.2.2	Accessories.....	3 6
6.3	Relocation	3 7
6.4	Keeping Product	3 7
7.0	Troubleshooting	3 8
7.1	Electrical.....	3 8
7.2	Trouble during Operation	3 9
8.0	Accessories	4 0
9.0	Appendix.....	4 2
9.1	Technical Specification.....	4 2
9.2	Disposing of the Product	4 3
9.3	Warranty.....	4 3
9.3.1	Warranty period.....	4 3
9.3.2	Technical Service Contact Points.....	4 3
9.3.3	Certificate of Product Warranty	4 4

1.0 Safety

1.1 Operation Manual

This manual contains important safety and operation information. You must carefully read, understand, and follow all the instructions in this manual prior to operating this instrument. Keep this manual in a safe place nearby for reference and make it easily available to all users.

1.2 WARNING/CAUTION/NOTICE Alerts

- (1) This manual highlights DANGER/WARNING/CAUTION/NOTICE alerts to prevent injury or property damage and also to achieve optimum performance of your instrument.
- (2) These alerts are classified into four types in this manual depending on the importance and the risk levels as described below:

Symbols	Meaning
	Indicates a hazardous situation which, if not avoided, will result in death or serious injury.
	Ignoring this warning could cause serious injury or even death.
	Ignoring this caution could cause injury or property damage.
	Ignoring this notice could cause operational problems.

1.3 Exemption for Responsibility

- (1) The claim which is out of the quality guarantee published by the Manufacturer is out of Manufacturer's responsibility.
- (2) The damage which is from unexpected fault or damage of user by Acts of God is out of Manufacturer's responsibility.

1.4 Warning Statement



Case of explosive and flammable chemicals, you must use with sufficient safety countermeasure.
In accordance with experiment, you should install safety devices and should follow suitable regulations in your laboratory.

Do not install the product in the place that the gas could leak out. Do not use in the place that has the industrial oil smoke and the metallic dust. It causes fire or electric shock.

Do not use the machine near to places where explosion can be happened due to organic evaporating gases.

Explosive materials: Acid, Esther, Nitro compound

Inflammable materials: salt peroxides, inorganic peroxide, salt acids.

Do not use the machine at places where moisture is high and flooding can be happened.

Check electrical requirements described in this operation manual or on the ID plate of this instrument before use.

Connect this instrument to a dedicated power outlet nearby.

Make sure to connect this instrument only to properly grounded power outlets to protect you and your instrument.


Do not ground to gas pipes or water pipes.

Please unplug when there are strange sound, smell and smoke from the product. Stop operating and request a service.

Do not assemble, repair, modify on your own.

The product may not work well and electric shock in the efficiency of the product. Also you cannot get after service by warranty regulation.

1.5 Caution Statement



Do not touch block immediately, even power off. Because the block might be have Residual heat. It may be cause burns from high temperature of block, seemingly invisible.

Do not forget plugging off, after product main switch off.
It is safety regulations for the next user.

Do not put heavy things on the power line. Do not put the machine on the line.
It may take off the wire coating and cause the electric shock or fire.

Do not touch the machine with wet hands.
It may cause electric shock or injuries.

Do not pour or insert any flammable materials into the product.

Do not pour water or put liquid on the top of the product when cleaning.
Please intercept the main power immediately and request the service when water may be in the product.

Do not let the product take any strong shock or vibration.
It causes abnormal operation or trouble. It may deteriorate the ability of the product and you may not obtain correct results.

Do not install the instrument near strong electric field exposed environment.

Please caution, the pace maker or magnetic recording instrument might be influenced by our instruments and magnetic stick

Do not sprinkle insecticide or flammable spray on the product. Use smooth cloths.
Cleaning with solvent can cause fire and deformity.

Please power off while product cleaning.
It may cause the electric shock or fire.

2.0 Functional Description

2.1 Introduction

CCB-350 is designed for high heat conductivity and fast heating rate by peltier module with 200W Heater.

General application follows that:

- Enzyme reaction
- Enzyme kinetics
- Immunoprecipitation kinase assays
- Sample incubation and reaction
- RNA transcription
- DNA analysis
- Genetic analysis – extraction of DNA, RNA and further sample preparation

2.2 Feature

2.2.1 Excellent Performance

- Precision accuracy is ensured (0.1°C) by its PID controller.
- Operation control range :
Available to use heating and cooling with provided wide temperature control range : 20°C below ambient ~ 95°C
- Using a peltier module offers fast temperature approaching.
- Users can conveniently set timer operation (1min ~ 99hour 59min). Two kinds of timers are provided for proper uses. Also, the timer can be check remaining time and resting. [Refer to 4.3 Timer Modes]
- Provide program modes that users can schedule as their experiment protocols.
 - Up to 10 programs allowed for memory storage. And each program can be set by maximum 10 steps temperature control functions.


2.2.2 Safety

- Designed by Water-proof structure that minimizes influx of reagents or solution.
- Designed by threefold safety system to cut off main power ; overheating protection for Heater, overheating protection for circuit, and over-current protection circuit structure.
- Polypropylene (PP) material covers, decreases the chance of accidental injuries.


2.2.3 Ease of Use

- Available to use 0.5mℓ, 1.5mℓ, 50mℓ etc. tube block. [Refer to 8.0 Accessories]
- Easy to check a set temperature value and operating conditions through VFD (Vacuum Fluorescent Display).
- Convenient to control the product by Touch Button and Dial Knob.
- The main body and control panel is made of polypropylene (PP). Polypropylene (PP) that is resistant to chemicals and is easy to clean.

2.3 Construction



- (1) VFD (Vacuum Fluorescent Display) : To display operating status.
- (2) Touch Button : To choose temperature and time
- (3) Dial Knob : To set a temperature and time
- (4) Lid : User can indicate sample state during the heating, and distribute uniform temperature.
- (5) 1.5mℓ tube block(Optional) : 0.5mℓ, 15mℓ, 50mℓ, etc. tube block can be used. [Refer to 8.0 Accessories]
- (6) Power Switch: Power should be ON/OFF.



- (7) Socket: Connecting the power cable into the socket
- (8) Extractor : Used when block exchange [Refer to 4.5]


3.0 Installation

3.1 Unpacking Package

- (1) Check to see if there is any damage in the instrument package before unpacking.
- (2) Unpack the instrument carefully.
- (3) Inspect to see that the instrument is not damaged during transportation.

3.2 Checking Instrument Components

- (1) Please check the instrument components supplied in the package after unpacking.
- (2) If a noticeable or an omission is found, immediately notify your local Jeio Tech dealer's Service Department.

Item	Figure	Quantity	Description
Main body		1	-
Extractor		2	-
Power cord		1	-
Operation Manual		1	-

3.3 Installing Environment

It should be installed in suitable environment as below.



Avoid direct sunlight.



Room temperature should be 5°C ~ 30°C.



Relative Humidity (RH%) should be less 80%.



Altitude should be less than 2,000m.

3.4 Location Conditions

Place the instrument far from the other instruments and keep the proper distance (normally more than 30 cm).



WARNING

- Place and install the instrument on a stable fireproof surface with non-slip and non-moisture and avoid direct sunlight & heat.


3.5 Checking Points

- It should be used on a flat working table with safety facility.
- It should not be used in the places where a combustible gas leak might occur.
- It should not be used in high electric field environments.
- It should not be used in places in danger of electric leakage, water leakage, and submersion.
- It should not be used where there are industrial harmful gases or metal dusts.

3.6 Connecting to Main Power Supply

Connect the electric power to the instrument according to the following process.

- (1) Switch off the main power switch before connecting the power cable.
- (2) Connect the power cable to the socket of the main body and to the power supply.



WARNING

Electrical Shock Hazard



- Ensure that the instrument should be connected to an appropriate power supply in terms of voltage, phase and capacity.
- Should use a grounded power.
- Never use a forked socket, or a double-tapped socket.
- Failure to obey a safety warning will cause a drop in a line voltage, resulting in a loss of power and causing risk of fire by turning the cable.
- Do not handle or touch electric codes and devices with wet hands.
- Wrong power supply can cause serious damages to the instrument and body; even to death.

3.7 Instrument Start and Stop

1. Plug In
When you switch on, Fan starts working and display as below.




NOTICE


- Touch Buttons are displayed with green or red colored backlight. When a backlight color of a button is green, the button is valid so that it can sense user's touches to itself. When it is red, it cannot sense the touches of users.
- When the instrument is power on, Fan starts operating.
- Before displaying the temperature, the instrument needs start-up latency time shortly.

2. Run/Stop


You can operate or stop the instrument by pressing START/STOP.




[Display when Heater is on]



[Display when Cooler is on]



[Display when Heater and Cooler are on]




[Display when Heater and Cooler stop]

NOTICE

- If you set temperature less 50°C, Cooler and Heater work at the same time.
 - If you set temperature over 50°C, Only Heater works.
 - There might be some noises when Fan or Cooler works.
-

4.0 Operation

4.1 Control Panel and Function



- (1) VFD (Vacuum Fluorescent Display) : You can check the setting temperature and current state of operation.

A	<i>Rem'n</i>	Remaining time during operation
B	RUN	Indicate the instrument is operating.
C	STOP	The instrument is stopped.
D	●	Indicate Heater is working.
E	⚙️	Indicate Cooler is working.
F	<i>PV</i>	Process Value (Current value)
G	<i>SV</i>	Set Value (Target value)
H	<i>Temp.</i>	Indicate temperature is now displayed.
I	<i>Time</i>	Indicate time(elapsed time or remaining time) is now displayed or timer is operating.

- (2) Touch Buttons


A		START/STOP	Start and stop the instrument. Go back one step when setting parameters for operation.
B		TEMP	Set temperature. Go into Offset setting.
C		PROGRAM	Step in the program mode.
D		TIMER	Set timer function. Check remaining time or elapsed time.

- (3) Dial Knob : It can be used when setting temperature, timer and program parameters.


4.2 General Modes

4.2.1 Setting Temperature

- (1) Press TEMP when the instrument is on standby




- (2) You can set a target temperature by adjusting Dial Knob with 0.1℃ resolution(e.g., change the SV from 23.5℃ to 60.0℃).




At any time, you can be out of temperature set mode by pressing START/STOP.

- (3) Pushing Dial Knob, you can save the target temperature as a temperature SV(Set Value).



- (4) Then you can operate it by pressing START/STOP.



NOTICE


- To step back from temperature setting, press START/STOP. The instrument returns to standby.
- You can also change the temperature SV when the instrument is in operation. [Refer to 4.2.2 Changing Temperature in Operation]
- The instrument keeps the temperature after the temperature PV reaches to its SV by repeating Heater and Cooler on and off.
- SV of temperature is updated to the last set value. This will not be initialized even after Power Switch off and on.
- You can set Offset by pressing TEMP during operation.

CAUTION

- The instrument and its accessories can be hot even though the Power Switch is off.

4.2.1-1 Changing Temperature in Operation

It is possible to change a target temperature by pressing TEMP during operation. Change the temperature SV by adjusting Dial Knob and you can start it by pushing Dial Knob.




NOTICE


- You can escape from the temperature changing process by pressing START/STOP or leaving it without additional input for 10 seconds.

4.2.1-2 Checking Elapsed Time in Operation

Elapsed time for operation can be checked by pressing TIMER once when the instrument is in operation.



To step back from the elapsed time check, press START/STOP or leave the instrument for 10 seconds.



4.2.1-3 Stopping Temperature Control Operation

Press START/STOP during General Mode operation. The instrument will stop operation.

CAUTION

- The instrument and its accessories can be hot even though the Power Switch is off.

4.3 Timer Modes

This instrument provides two types of timers (T1 and T2 timer). The timers are different in a operating condition and point of time countdown.

T1 timer	T2 timer
Set when heating control is not operating	Set when heating control is operating
Start counting down the timer just after it reaches temperature	Start counting down the timer immediately

NOTICE

- Timer can be set from 1min~99h 59min.
- You can be out of timer setting in Timer mode by pressing START/STOP repeatedly.
- T1 timer starts counting down when temperature reaches to SV and “Time” in VFD blinks.
- T2 timer starts counting down when set the timer, immediately and “Time” in VFD blinks.
- You can check elapsed time by pressing TIMER once and get into T2 timer setting by pressing TIMER twice during temperature control operation.
- After a timer counting is finished, the Cooler starts to make the instrument temperature converge to the room temperature.
If the user confirms the end of the timer operation by using Touch Buttons or Dial Knob, then the Cooler does not operate and the instrument returns to standby state.

4.3.1 Setting T1 Timer

T1 timer can be set only when the instrument stops temperature control.


After setting T1 timer, the temperature control starts operating to make the temperature PV the same to the temperature SV.

When its temperature reaches to the temperature SV, the timer starts to count down.


- (1) Press TIMER when the instrument stops.




- (2) Input Hour, Minutes by using Dial Knob.
(Timer range: 1min~ 99hour 59min)



- (3) Input a target temperature by using Dial Knob and save it by pushing Dial Knob.
In case of do not changes set temperature, just push Dial Knob.




- (4) Select whether start operating the instrument or not by using Dial Knob.



When it reaches to the set temperature, the timer starts with a sound alarm.

- (5) When the timer operation ends, the instrument generate a sound alarm with a display as follows. Confirm the end of the timer operation by using Touch Buttons or Dial Knob.



NOTICE


- If you press Dial Knob in “RUN NO”, all inputs for T1 timer are canceled and the instrument returns to the standby.
- Every set target time (time SV) is memorized as an initial value of T1 timer. You can see the last set time SV when setting a new time SV, as an initial value.
- During the timer operation, you can stop the instrument by pressing START/STOP. [Refer to 4.3.4 Stopping Timer during Operation]
- During timer setting, the instrument returns to standby if you leave it for about 10 seconds without saving the time value.

4.3.2 Setting T2 Timer


T2 timer is for setting a timer during temperature control operation. It also requires SVs for temperature and time.

In T2 timer mode, regardless of reaching the temperature SV, the timer starts to count down just after the timer setting.


- (1) Press TIMER twice during temperature control operation.




- (2) Select a desired time by using Dial Knob in the order of hour, minute.
You can set the timer from 1min to 99hour 59min.



(3) Input a target temperature by using Dial Knob and save it by pushing Dial Knob.



(4) When the timer operation ends, the instrument generates a sound alarm with a display as follows. Confirm the end of the timer operation by using Touch Buttons or Dial Knob.




NOTICE

- In case of T2 timer, the timer starts to count down just after the timer setting, regardless of reaching the temperature SV.

4.3.3 Checking Remaining Time during Timer Operation

Remaining time for timer operation can be checked by pressing TIMER once when the instrument is in operation.




[Remaining time]

During the checking process, you can escape from the process if you press START/STOP or leave the instrument without additional input for 10 seconds.

4.3.4 Stopping Timer during Operation

You can stop the operation by pressing START/STOP.



Confirm the end of the timer operation by using Touch Buttons or Dial Knob.

CAUTION


- The instrument and its accessories can be hot even though the Power Switch is off.

4.4 Program Mode

This instrument provides 10 user-programmable programs. Each program consists of steps(max.10 steps) that have parameters including temperature SV and time SV(step duration).


It also provides a looping function that repeats a program as many times as user selected (range : 1~99, infinity).

[Diagram of Program Mode]



ABBREVIATIONS	DESCRIPTIONS
PROG(PR)	Program
EXEC	Run a program
VIEW	View configurations of a program
EDIT	Edit configurations of a program
CLER	Clear all programs or a program.

[Temperature and shaking speed control in Program Mode]




NOTICE


- A time SV for a step starts to be counted just after the instrument temperature reaches to the temperature SV of the step.
- If the time SV for a step is set as “00:00”, then the step will be skipped and the next step will proceed.
- It is possible to repeat a program as user selected times from 1 to 99, or infinite.
- You can stop the operation of a program mode by pressing START/STOP. [Refer to 4.4.5 Stopping Program]
- User can repeat the saving program from 1 to 99 times or infinitely.

4.4.1 Starting User Program

- (1) Press the PROGRAM when the instrument is on standby.



- (2) Select one of user programs by turning Dial Knob to the right side.





- (3) Select a program execution “PR n EXEC” by pushing Dial Knob (n:program number).




(4) Select the number of loops for program repeats by turning Dial Knob and operate it by pushing Dial Knob.

- Turn Dial Knob to the right : You can select the number of repeats from 1 to 99.
- Turn Dial Knob to the left : Repeat the program infinitely.




4.4.2 Display during Program Operation

During program mode operation, the instrument displays its operating status, “program number and current step” → “remaining time of current step” → “the current number of repeats in infinite loop” or “the current number of repeats and total number repeats” → “process value and set value for temperature”.



[Program 2 and step 3] → [remaining time : 5min and step 3] → [10th iteration of infinite loop] → [temperature – PV : 60°C, SV : 60°C]

In case of ramping periods between two different steps, the display for remaining time is replaced to alternative displays indicating that the instrument is on temperature ramping.




[Program 2 and step 3] → [ramping to step 3 temperature] → [1st iteration of 5 total repeats] → [temperature-PV : 23.5 °C, SV : 60 °C]

4.4.3 Editing User Program


- (1) Press PROGRAM when the instrument is on standby.




- (2) Select a program you want to edit by turning and pushing Dial Knob.



- (3) Select a program execution "PR n EDIT" by turning and pushing Dial Knob as below (*n*:program number).





- (4) Enter the number of steps for the program. You can set up 10 steps for a program.




- (5) The program is reset and generates new steps as many as step SV. Then you can set SVs for temperature, time and shaking RPM for each step by using Dial Knob.



- (6) Enter the target temperature and time for the step, and then save the entered value (SVs) as below.






If you skip the saving process, all entered inputs are not saved.

By applying (5) and (6) for every step, you can complete program editing.

(7) After saving all generated steps, push START/STOP to escape program editing.




NOTICE


- If you want to write a new program on a program slot, set step size of the program slot. The step size setting includes clearing existing program and initializes a new program. [Refer to 4.4.7-2 Reset Individual Program]
- Just after you set the step size for a program, each step has 4°C for temperature and 00:00 for time as initial values. This instrument is supposed to skip steps that have 00:00 for time. So, non-edited steps will be skipped in program execution.
- Temperature setting range for this instrument is 4°C~95°C, Recommended temperature control range is from 20°C below ambient to 95°C.
- When program operation ends, Cooler automatically operates until the temperature goes down to room temperature. If you confirm the end of the program by using Touch Button or Dial Knob, Fan stops and the instrument become standby.

4.4.4 Viewing Program Setting


- (1) Press PROGRAM when the instrument is on standby.




- (2) Select a program you want to view by turning and pushing Dial Knob.




- (3) Select a program execution "PR n View" by turning and pushing Dial Knob as below (n:program number).




- (4) Turn Dial Knob to find a specific step you want to view. Push Dial Knob to get into the step.



And then you can view the SV of temperature and time for the step as below.




(5) After viewing the program setting, move out by using START/STOP.




4.4.5 Stopping Program

By pressing START/STOP, you can stop program operation.



4.4.6 Confirming Program End

"END PR *n*" is displayed with sound alarm at the end of a program (*n*:program number). Please check the program end by pressing START/STOP.




4.4.7 Resetting Program

1. Reset All Programs

(1) Press PROGRAM at standby.




(2) Turn Dial Knob to the right end and select "PROG CLER".




All programs are initialized. It takes some time for resetting all programs.

2. Reset Individual User Program

(1) Press PROGRAM at standby.



(2) Select a target program that needs to be deleted.




[In case of resetting program 2]

(3) Turn Dial Knob to the right and push “PR *n* EDIT” (*n*:program number).



(4) Turn Dial Knob to the right end and press Dial Knob at “STEP CLER”.




All previous steps of the program are deleted.

NOTICE


- Setting a step size (“STEP SV”) for a program includes resetting the previous program and generating new steps as many as the set step size.
-

4.5 How to Replace a Block


- (1) Remove a extractor from the back of the instrument by turning the extractor in a counter-clockwise direction.



- (2) Fix the detached extractor to a hole in the middle of a block, by turning it clockwise as below.



- (3) Detach the block by lifting it up from the instrument in a direction of an arrow shown as shown below.



(4) Insert a new block into the instrument through using the extractor in the same way as above.

CAUTION

- Before replacing a block, please detach tubes, vials, microplate etc. from the block.
- You should be careful in handling the instrument. Blocks and main body of the instrument can be hot even after the Power Switch is off.
- Make sure that blocks and an extractor are inserted correctly.

4.6 Offset


The temperature shown on the VFD is measured by a temperature sensor inside the instrument. However, this temperature can be different from the temperature of your own thermometer which you may use as a standard for your specific applications. If needed, you can offset such temperature differences within the range of $\pm 50^{\circ}\text{C}$ at 0.1°C interval.

The temperature offset setting procedures are as follows:


(1) Turn the power on.




(2) Press TEMP until the following temperature offsetting screen appears with audible alarm.




(3) Select the offset value by turning the Dial Knob appropriately and, when selected, save it by pushing Dial Knob. When properly saved, the save confirmation screen will appear as shown below:



- (4) After confirm saving offset, setting a SV temperature is automatically displayed as below. Finally, it returns to standby state with PV that offset is applied to and the changed SV



If you press START/STOP or leave the instrument for about 15 seconds in this step, then it returns to standby and you can see that PV is affected by the offset.



NOTICE

- You can except the offset by pressing the START/STOP during offset input.
- During operation, you can set offset by pressing TEMP.

5.0 Safety Device

- (1) The overheat protection device for Block.
The power of this instrument automatically cuts off when the block's temperature overheats more than 115°C , to protect the instrument.
- (2) The circuit protection device.
The power automatically switches off when the circuit's temperature overheats in a certain degree which will activate this device to protect the circuit inside of the instrument.
- (3) The overcurrent protection device.
The power automatically switches off when the currency flows more than the given currency value which will activate this device to protect the instrument.

NOTICE

- The instrument should be used after fully cooling it down when the instrument is switched off by these kinds of the protection devices.
-

6.0 Maintenance

6.1 Periodic Maintenance

5tclassifications	A period of checking time	
	daily	weekly
Power cord		
- The conditions of connection for power supply and an adaptor	•	
- The presence of power supply and an adaptor contact wetting, and cable peeling off, and out of contact.	•	
Product surface cleaning		•
Block cleaning condition	•	
Controller function checking	•	
Check accessory attachments to the instrument are tight.	•	

6.2 Cleaning Product

6.2.1 Main Body

Remove a contaminant by cleaning the instrument frequently with a soft cloth before and after using, otherwise it cannot be readily wiped out for a long time Keep the instrument cleans always without any contaminant.

CAUTION

- Do not put under the water.
 - Do not damage inside of accessories and system. Please caution.
 - Do not touch to product from high concentration of nitric acid, sulfuric acid, sodium hydroxide, acetone, benzene, phenol, toluene, chloroform, cresol, acetic acid series, and chlorine series corrosive solvent.
 - Please separate power cord from body, if do not use.
 - Do not use chlorine bleach, ammonia-based cleaners, abrasives, ammonia, or metal scouring pads when cleaning.
-

6.2.2 Accessories

Remove a contaminant by cleaning the instrument frequently with a soft cloth before and after using, otherwise it cannot be readily wiped out for a long time Keep the instrument cleans always without any contaminant.

6.3 Relocation

- (1) Disconnect the power cord from the power outlet.
- (2) Pack the instrument and its accessories into the original packaging or any other suitable container before moving.

 CAUTION

- Pay attention to avoid mechanical shock or vibration while moving the instrument. Damages caused by mechanical shock or vibration may result in injury or fire.
-

6.4 Keeping Product

- (1) Unplug the instrument from the main power.
- (2) Clean the instrument with a soft cloth neatly.
- (3) Store in a dry place after packing.

7.0 Troubleshooting











7.1 Electrical




Trouble	Causes	Solution
The instrument does not turn on	Incorrect electric power	Compare power source and voltage on the ID plate and make sure they are the same. ID plate is found on the back of instrument.
	Power failure or circuit breaker shuts down	Find out the causes of power failure and recovery.
	Main plug not seated properly.	Check the electrical cord connection at the instrument to ensure it is fully seated.
	Check the electrical cord connection at the instrument to ensure it is fully seated.	If the socket / plug / main power line are cut, request service.
	PCB has damaged by reagent	Request service.
Room circuit breaker trips often when the instrument is turned on or running	Too many plugs connect at the same time	1. Check the circuit breaker size along with the voltage and current supplied to it. 2. Check that several similar instruments are inserted together, if so you should not use overly.
	Product internal circuit problem	Request service
No VFD	Power failure	Find out the causes of power failure and recovery.
	Main plug not seated properly.	Check the electrical cord connection at the instrument to ensure it is fully seated
Button doesn't operate well	Power failure	Find out the causes of power failure and recovery.
	Button switch has damaged	Request service.

7.2 Trouble during Operation

Symptom	Cause	Action
Block is not heating up	START/STOP is not pressed.	Press START/STOP on the control panel.
	Set temperature is lower than present temperature	Check set temperature and adjust it properly.
	Heater failure	Request service
	Circuit protection device cut off the power.	Take off the cord for cooling down the instrument and re-operate it.
	Product internal circuit problem	Request service
The Dial Knob isn't operation correctly	Dial Knob or circuit problem.	1. Pull out the knob from the instrument and replace it again. 2. Request service.
Error message on the display	Product internal circuit problem	Request service
the VFD lamp is not operating	Product internal circuit problem	Request service

8.0 Accessories

Designation			Order No.	Description (Dimension : W x D x H, mm)	
Blocks	1.5mℓ X 30		00CHB00000045	98 X 76.5 X 41	You can close or open the instrument cover, when using these blocks.
	0.5mℓ X 48		00CHB00000029	98 X 76.5 X 41	
	48-well		00CHB00000046	98 X 76.5 X 41	
	15mℓ X 15		00CHB00000031	98 X 76.5 X 51	You cannot close the instrument cover, when using these blocks.
	50mℓ X 6S		00CHB00000032	98 X 76.5 X 51	
	50mℓ X 6T		00CHB00000033	98 X 76.5 X 87	
	Ø10 X 35		00CHB00000034	98 X 76.5 X 51	
	Ø12 X 24		00CHB00000035	98 X 76.5 X 51	
	Ø13 X 24		00CHB00000036	98 X 76.5 X 51	
	Ø15 X 20		00CHB00000027	98 X 76.5 X 51	

	Ø16 X 16		00CHB0000037	98 X 76.5 X 51	
	Ø18 X 12		00CHB0000038	98 X 76.5 X 51	
	Ø20 X 12		00CHB0000039	98 X 76.5 X 51	

9.0 Appendix

9.1 Technical Specification

Heating & cooling block		CCB-350
Temp.	Control range (°C/°F)	Amb. -20 to 95 / Amb. -36 to 203 (PID feedback)
	Setting range (°C/°F)	4 to 95 / 39.2 to 203, 0.1°C resolution
	Uniformity(±°C/°F)	0.06 / 0.11 (at 37°C)
		0.48 / 0.86 (from 37 to 95°C)
	Stability(±°C/°F)	0.15 / 0.27 (at 37°C)
		0.35 / 0.63(from 37 to 95°C)
	Heater output, max. (W)	200
	Peltier output, max. (W)	160
	Heating rate (°C/ min)	Approx. 5.5
	Cooling rate (°C/ min)	Approx. 0.9
	Function	Offset
Program		10 memories, 10 steps/memory
Timer		1min to 99 hr 59 min
Safety device		Over temperature protection Over current detection
Control panel		VFD(Vacuum Fluorescent Display), 4 Touch Buttons, Dial Knob
Material	External	PP, PC, Powder coated steel
	Block	Black anodized Aluminum
Dimension	Internal(mm/inch)	99X77.5 X36 / 3.9X3.1 X1.4
	Overall(mm/inch)	249X330X168 / 9.8X13 X6.6
Net weight (kg/lbs)		5.0 / 11.0
Electrical requirements (120V/60Hz,A)		4
Electrical requirements (230V/50/60Hz,A)		2
Exchangeable blocks		1.5mℓX30, 0.5mℓX48, 48-well, 15mℓX15, 50mℓX6S, 50mℓX6T, Ø10X35, Ø12X24, Ø13X24, Ø15X20, Ø16X16, Ø18X12, Ø20X12

※ Specifications can be changed without prior notice for quality upgrade.

※ Permissible ambient condition: 2~60°C, Relative humidity up to 80%.

9.2 Disposing of the Product

Before disposing of the shaker or any of its components:



1. The instrument should be cleaned and decontaminated to protect workers servicing the instrument, the environment or the public purchasing surplus instrument because the shaker can potentially be contaminated with biological material, chemicals or radioisotopes. Check with your institution or laboratory for individual policies and procedures for disposal of laboratory instrument.
2. Please contact your local governing body for regulations regarding disposal of electrical, electronic, metal (brass, aluminum, steel and stainless steel), refrigeration and rubber components. Jeio Tech recommends the user find a local scavenger or laboratory instrument recycler to properly dispose of the instrument and its components.

9.3 Warranty

9.3.1 Warranty period

- (1) If the product fails during normal and proper use, the warranty period for manufacturing liability shall be two years from the date of delivery.
- (2) If you request a repair, please check the items below and let us know so that you can get a quicker and more accurate repair.

- | |
|---|
| <ul style="list-style-type: none">• Date of purchase:• Customer name / address / phone number / E-mail:• Fault status• Model• Serial number |
|---|

9.3.2 Technical Service Contact Points

Technical Services

Address : 153 (Yongsan-dong), Techno 2-ro Yuseong-gu, Daejeon-city, 34025, Republic of Korea

Tel : +82 (0)2 -2627-3824

Website : <https://www.jeiotech.com/eng/>

E-mail : overseas@jeiotech.com

9.3.3 Certificate of Product Warranty

This is a guide to the product quality warranty and service terms provided by JEIOTECH Co., Ltd

All products are guaranteed to operate normally when installed under proper conditions, as specified in the user manual for each product, and used according to the intended purposes and permissible operating conditions.

The warranty period for product quality starts from the delivery date and is indicated in the catalog provided for each product or in the purchase order agreed upon with the customer. If a product malfunctions within the warranty period, JEIOTECH Co., Ltd. provides free repair services. However, among the transportation costs incurred during the repair process, any special costs resulting from the customer's site conditions (e.g., door or wall modifications, use of special handling equipment) are not covered under the warranty. Furthermore, JEIOTECH's liability under this warranty is limited to replacing the product or refunding the original purchase price at JEIOTECH's sole discretion if reasonable attempts to repair or replace the product fail.

Parts replaced during free repairs within the warranty period become the property of JEIOTECH Co., Ltd. Additionally, the warranty period is not extended due to free repairs. JEIOTECH is not responsible for any losses incurred by the customer due to delays in repairs caused by issues such as part shortages or shipping delays.

The following cases will be handled as paid repairs, even within the warranty period:

1. Non-defective cases

- * Requests to resolve issues arising from improper operating environments or abnormal installations.
- * Reinstallation requests or connections with external equipment other than the initial installation.
- * Requests for usage instructions or simple adjustments that do not require disassembly.
- * Inspections of non-defective products.
- * Cleaning or debris removal service requests.

2. Failures caused by user errors:

- * Malfunctions due to usage not in accordance with the user manual.
- * Issues resulting from improper operating environments or abnormal installations.
- * Malfunctions caused by incompatible chemical usage.
- * Malfunctions caused by irregular or improper electrical or utility supplies.
- * Damage due to product movement, storage, external impact, or dropping by the customer.
- * Malfunctions resulting from unauthorized modifications or repairs without prior written approval from JEIOTECH Co., Ltd.
- * Malfunctions caused by using consumables or optional components not specified by JEIOTECH Co., Ltd.
- * Repairs performed by personnel other than JEIOTECH or its authorized service agents.

3. Other cases

- * When the product model or serial number has been damaged, removed, manipulated, or altered.
 - * Malfunctions caused by fire, flooding, or other natural disasters.
 - * Failure due to consumable parts reaching the end of their lifespan (e.g., lamps, filters, fuses, gaskets, packings, hoses, heat transfer fluids).
 - * Natural wear and tear (e.g., discoloration, screen printing degradation).
 - * Damage or malfunction of product accessories (e.g., power cables, data cables, glass components) or options.
 - * Malfunctions caused by errors in computers, computer components, or connected network systems.
- Before JEIOTECH performs free or paid repairs or accepts the return of a product, the customer must ensure the product has been decontaminated of chemical and biological hazards.
 - Replacement parts used during repairs may differ from the original but will function correctly and be equivalent in performance.
 - JEIOTECH does not bear any responsibility for damage to stored samples, data loss, or any other direct or indirect damages caused by product malfunctions, regardless of the warranty period.
 - JEIOTECH does not guarantee issues related to computers, computer components, or network systems connected to the product, regardless of the warranty period.
 - This warranty is the sole and exclusive warranty provided. Any implied warranties imposed by applicable laws are limited to the warranty period stated herein. This warranty cannot be modified by any verbal advice, statements, or documents on behalf of JEIOTECH unless explicitly labeled as a "Warranty Amendment" and signed by an authorized representative of JEIOTECH.

- This warranty applies only to products purchased for direct use by the customer. It does not cover products purchased for resale purposes without prior written approval from JEIOTECH.
- This warranty does not apply to consumables and is non-transferable.

- The contents of this manual can be changed or upgraded without prior notice.
- The copyright of this manual is reserved by Jeio Tech.